




Hemax 330

Auto Hematology Analyzer



20 Parameters & 3 Histograms
60 samples per hour
9.8 μ L samples volume
Operation with touch screen

Features

- 20 Parameters+3 Histograms for WBC,RBC and PLT
- 3-Part differentiation of WBC
- 3 Counting mode: whole blood, Pre-diluted Blood, Peripheral Blood
- LCD display with Touch Screen and Mouse Operation
- Throughput: 60 samples per hour
- Built-in thermal printer, optional external printer
- External AC adapter avoids electronic noises, enhancing accuracy
- Up to 50,000 sample results(including histograms) stored
- Multi-language



Hemax 330

Technical Specifications

Parameters:

WBC, RBC, HGB, MCHC, MCH, MCV, HCT RDW-CV, RDW-SD, PLT, MPV, PDW, PCT, P-LCR, LYM#, MID#, GRA#, LYM%, MID%, GRA%, 3 Histograms for WBC, RBC and PLT

Throughput: Up to 60 samples per hour

Carryover: WBC/RBC/HGB < 0.5%, PLT < 1.0%

Accuracy (CV):

- WBC \leq 2.0%
- RBC \leq 1.5%
- HGB \leq 1.5%
- MCV \leq 0.4%
- PLT \leq 4%

Principle: Electrical resistance for counting WBC, RBC and PLT

Sample Volume: Whole Blood 9.8 μ L, Pre-diluted Blood 20 μ L, Peripheral Blood 9.8 μ L

Operating Environment: 15°C-35°C, Humidity \leq 80%

Alarms: Error Messages

Input/Output: RS-232, USB, LAN, Keyboard and Mouse Interface

Printout: Thermal printer, 55mm width paper, Various printout formats, External printer optional

Power Requirement: a.c.110V-220V, 50/60Hz

Dimension: 478mm(L)X330mm(W)X395mm(H)

Weight: 23kg



B&E SCIENTIFIC INC

Add:804 Thorndale Ave, Bensenville IL 60106
Tel:+1 630 974 5159
Fax:+1 630 974 5160
Email:info@besic.com
Web:www.besic.com

Your Local Distributor is:



للطلب و الاستفسار

أوميغا لاب للتجهيزات المعملية

www.ome-ga.com

01066400909

01117673062

01066404446

01019026268

01066400922

01277233260

أوميغا للتجهيزات المعملية

